KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

# SCHEDULE OF ACCREDITATION

## KENYA PIPELINE COMPANY LIMITED

### ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/037**

**PERMANENT ADDRESS OF CAB**

KenPipe Plaza, Sekondi Road,

Off Nanyuki Road, Industrial Area -Nairobi

P. O. BOX 73442-00200

NAIROBI, KENYA

Tel: 020 – 2606500-4

Email: cc@kpc.co.ke

Website: www.kpc.co.ke

**Laboratory Contact Person**

Amos Mugira

Expiry Date: 07-July-2027

Approved by: _____ Date: 29-SEPT-2025

~~KENAS CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

# SCHEDULE OF ACCREDITATION – KENYA PIPELINE COMPANY

| No. | Testing Field | Test Object/ Matrix | Test Parameter | Test Method | Type of Test | Measurement Range | Test Location |
|---|---|---|---|---|---|---|---|
| 1. | Chemistry | Gasoline / MSP | Distillation | ASTM D86 | Physical Chemistry | Max 210 °C | Nairobi (PS10) Mombasa (PS14) |
| | | Gas Oil | Distillation | | | ≤ 400 °C | Mombasa (PS1) Nairobi (PS10) |
| | | Jet A-1, Illuminating Kerosene | Distillation | | | ≤ 300 °C | Mombasa (PS1) |
| | | Jet A-1, Illuminating Kerosene, Dual Purpose Kerosene | Distillation | | | ≤ 300 °C | Mombasa (PS14) Eldoret (PS27) Kisumu (PS28) Nairobi (PS10) |
| | | | Density at 20 °C | ASTM D1298 | | 0.775 – 0.840 Kg/L | Nairobi (PS10) Kisumu (PS28) Eldoret (PS27) Nakuru (PS25) |
| | | | Density at 15°C and at 20 °C | ASTM D4052 | | 0.775 – 0.840 Kg/L | Nairobi (PS10) Kisumu (PS28) Mombasa (PS1 & PS14) Eldoret (PS27) Nakuru (PS25) |
| | | | Abel Flash point | IP 170 | | 9 - 70°C | Nairobi (PS10) Kisumu (PS28) Mombasa (PS1 & PS14) Eldoret (PS27) Nakuru (PS25) |
| | | | Automatic Colour Saybolt | ASTM D6045 | | +30 to -16 Colour Units | Nairobi (PS10) Kisumu (PS28) Mombasa (PS14) Eldoret (PS27) |
| | | | Automatic smoke point | ASTM D1322 | | 15.0 – 30.0 mm | Nairobi (PS10) Mombasa (PS14) |
| | | | Automatic Freezing Point | ASTM D5972 | | Max - 47°C | Nairobi (PS10) |
| | | | Automatic Freezing Point | ASTM D7153 | | Max - 47°C | Mombasa (PS1 & PS14) Nairobi (PS10) Eldoret (PS27) Kisumu (PS28) |
| | | Gas Oil (AGO) | Flah point | ASTM D93A | | Min 60°C | Mombasa (PS1 & PS14) Nairobi (PS10) Eldoret (PS27) Kisumu (PS28) Nakuru (PS25) |
| | | | Density at 20 °C | ASTM D4052 | | 0.817 – 0.867 Kg/L | Mombasa (PS1 & PS14) Nairobi (PS10) |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| | | | | | | | Nakuru (PS25)<br>Eldoret (PS27)<br>Kisumu (PS28) |
| | | Gasoil (AGO) | Density at 20 °C | ASTM D1298 | | 0.817 – 0.867 Kg/L | Eldoret (PS27) |
| | | Gasoline (MSP) | Density at 20 °C | ASTM D4052 | | 0.716 – 0.771 Kg/L | Mombasa (PS1 & PS14)<br>Nairobi (PS10) |

## Accreditation History

| Type | Decision Date | Effective Date | Expiry |
|---|---|---|---|
| **Initial Assessment** | 08-JUL-2016 | 08-JUL-2016 | 07-JUL-2019 |
| **1st Re-accreditation** | 20-FEB-2019 | 08-JUL-2019 | 07-JUL-2023 |
| **2nd Re-accreditation** | 22-JUL-2023 | 08-JUL-2023 | 07-JUL-2027 |

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 02-Oct-2025

~~KENAS CEO~~/ Authorized Representative