KENAS
Delivering Trust
KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya
Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

NATIONAL WATER AND SEWERAGE CORPORATION

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/097**

PERMANENT ADDRESS OF LABORATORY

CENTRAL LABORATORY

PLOT 2, NAKASERO ROAD,

P. O. BOX 7053

KAMPALA, UGANDA.

Tel: +256 (0) 782269252
Email: Mohammed.babu@nwsc.co.ug
Website: www.nwsc.co.ug

Laboratory Contact Person:

Dr. Babu Mohammed

Technical Signatories

John Kennedy

Robinah Muheirwe

Expiry Date: 02-November-2029

Approved by:

Date: 03-November-2025

~~KENAS CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – NATIONAL WATER AND SEWERAGE CORPORATION-CENTRAL LABORATORY

No	Testing Field	Test Object / Matrix	Test Parameter	Test Method	Type of Test	Measurement Range	Test Location
1	Physical	Water/ Wastewater	pH	APHA – 4500-H	Potentiometric	1-14	Central Laboratories
2	Chemical	Water/ Wastewater	Electroconductivity	APHA-2510	Potentiometric	0-12ms/cm	Central Laboratories
3	Chemical	Water/ Wastewater	Turbidity	HI-88713-ISO	Physical	0-800NTU	Central Laboratories
4	Chemical	Water/ Wastewater	Alkalinity	APHA-23206	Titration	0-800mg/l	Central Laboratories
5	Microbiology	Water/ Wastewater	Escherichia coli	APHA-9221F	Enumeration	0-200cfu/ml	Central Laboratories
6	Microbiology	Water/ Wastewater	Feacal coliform	APHA-9221F	Enumeration	0-2000 cfu /ml	Central Laboratories

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	03-Nov-2025	03-Nov-2025	02-Nov-2029

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 03-November-2025

~~KENAS CEO~~/ Authorized Representative