KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

INSPECTORATE EAST AFRICA LIMITED

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/044**

PERMANENT ADDRESS OF LABORATORY

OFF MNAZI MOJA ROAD LIMITED MOMBASA, KENYA PLOT NO.409

P.O.BOX 42327-80100 GPO

MOMBASA, KENYA

Tel: +254 723 282 499, +254 714 323 930

Office No: 041 231 73 18, 041 222 8259

Email: admin@inspectorate-ea.com

Website: www.inspectorate-ea.com

Laboratory Contact Person:

Robert Eddie Ogwangi

Technical signatory

Ronald Ndeche

Kyalo Ilunga

Raphael Munyao

Expiry Date: 29-September 2030

Approved by: _____ Date: 27-August-2026

~~KENAS CEO~~/ Authorized Representative

KENAS Delivering Trust | KENYA ACCREDITATION SERVICE

P. O. Box 47400 – 00100 | TEL. +254 - 20 484 0000 | Nairobi, Kenya

Email: info@kenas.go.ke | Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – INSPECTORATE EAST AFRICA LIMITED

S/No	Testing Field	Test Object / Matrix	Test Parameter	Test Method	Type of Test	Measurement Range	Test Location
1.	Chemistry	Automotive Gas Oil and Jet A-1	Distillation	ASTM D 86	Volatility	Maximum 36 - 400°C	Inspectorate East Africa Limited Testing Laboratory
2.		Automotive Gas Oil	Density @ 20°C	ASTM D 1298	Gravimetric	0.6000 – 1.2500 Kg/L	
		Jet A-1	Density @ 15°C			775-840 Kg/m^3	
3.		Automotive Gas Oil	Flash Point, (Pensky Martens Closed Cup °C)	ASTM D 93	Volatility	19°C - 400°C	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	30-Sep-2019	30-Sep-2019	29-Sep-2022
1st Re-accreditation	01-Dec-2022	01-Dec-2022	29-Sep-2026
2nd Re-accreditation	27-Aug-2026	30-Sep-2026	29-Sep-2030

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past

Approved by: _____ Date: 27-August-2026

~~KENAS CEO~~/ Authorized Representative