



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya
Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION

VIVO ENERGY UGANDA

ISO/IEC 17025:2017

Testing Laboratory Number: **KENAS/TL/062**

PERMANENT ADDRESS OF CAB

PLOT 9/11, 7TH STREET, INDUSTRIAL AREA
P.O. BOX 7082, KAMPALA UGANDA.

Telephone (+256) 312210010/ 772754028

Email: Kennedy.Ggwanga@vivoenergy.com, Josephine.Nakato@vivoenergy.com

Laboratory Contact Person(s)

Ggwanga Kennedy
Nakato Josephine

Approved by: _____

KENAS CEO/Authorized Representative

Date: 21-December-2023



KENYA ACCREDITATION SERVICE

P. O. Box 47400-00100, TEL. +254-787-395679, +254-725-227640 Nairobi, Kenya

Email: info@kenyaaccreditation.org Web: www.kenas.go.ke

SCHEDULE OF ACCREDITATION – VIVO ENERGY UGANDA TESTING LABORATORY

No.	Testing Field	Type of Test	Test Method	Test Object/ Matrix	Test Parameter	Measurement Range	Test Location
1.	Chemistry	Volatility	ASTM D 86-18	Gasoline and Diesel	Distillation	400°C Max	VEU- Testing Laboratory
			ASTM D 93A-18 Method A	Diesel	Flash Point	19°C - 400°C	
		Gravimetric	ASTM D 1298-12b	Gasoline and Diesel	Density	0.775 - 0.840 Kg/m ³ @15°C	
			ASTM D 4052- 18a			0.68 - 0.97 Kg/m ³	

Accreditation History

Type	Decision Date	Effective Date	Expiry
Initial Accreditation	10-Mar-2020	10-Mar-2020	09-Mar-2024
1 st Re-accreditation	21-Dec-2023	10-Mar-2024	09-Mar-2028

This schedule is issued subject to the terms and conditions of KENAS Accreditation. It supersedes any other schedule(s) issued in the past.

Approved by:

KENAS CEO/Authorized Representative

Date: 21-December-2023